


Column : AppliChrom SugarSep-H V; 300mm x 8mm; P/N sasHV103008 Detection : RI
 Mobile Phase : H2O / 0.1% H2SO4 Temperature : 40C
 Flow Rate : 0.5 mL/min Pressure : -
 Note : Sample Test Mix /ApplM31HV / Organic Acids



Clarity Lite - 10.05.2023 14_11_03

	Reten. Time [min]	Compound Name	Asymmetry [-]	Efficiency [th.pl]	Eff/I [t.p./m]	Resolution [-]
1	12,65	Tartaric Acid	0,98	7940	26467	
2	13,50	Malic Acid	1,02	8395	27984	1,46
3	15,77	Succinic Acid	0,55	6024	20078	3,26
4	16,42	Lactic Acid	0,93	8241	27471	0,85
5	22,68	Propionic Acid	1,02	8822	29406	7,42